

# A124 : Filler compaction apparatus

**STANDARDS:** EN 1097-4 comparable to BS 812 | CNR N° 123 | NLT 177

The apparatus consists of:

- cylinder with an inside diameter of 25.4 mm;
- plunger freely sliding into the cylinder with max. lateral play of  $0.20 \pm 0.05$  mm;
- four columns and metallic base holding the whole.

To perform the test a measuring device (vernier caliper with 0.01 mm accuracy) is required: see accessory.

**Weight:** 4 kg approx.



## Accessories :

**A124-01** FILTER PAPER Ø 25 mm (pack of 100).

**V175-03** DIGITAL VERNIER CALIPER 0 - 150 mm x 0.01 mm sens.