

# E083 : Plunger penetration apparatus

**STANDARDS :** EN 413-2, EN 459-2, EN1015-4

Used to determine the consistency of fresh mortar, lime and masonry cement.  
The base is foreseen of a device to locate the test cup. The height of the drop can be accurately adjusted to 100 mm.

Supplied complete with test cup and tamper, both anodized aluminium made.

**Dimensions:** 200x200x700 mm

**Weight:** 8 Kg approx.



**E083**

BAG-MacBen BV  
Oostjachtpark 12,  
B-9100 Sint-Niklaas

T: +32 3 771 48 04 [info@macben.eu](mailto:info@macben.eu)  
[www.macben.eu](http://www.macben.eu)

BAG-MacBen  
Smederijstraat 2,  
NL-4814 DB Breda

T: +31 765 30 23 73 [sales@macben.eu](mailto:sales@macben.eu)  
[www.macben.eu](http://www.macben.eu)